



West End Grange
Epworth Water Gardens
West End Road
Epworth
Doncaster
DN9 1LA



Certificate No:

50018

Tel: 01427 875 867
Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name:

Address:

Postcode:

Contact Name:

Test Date:

Location:

Manufacturer:

Serial/Plant No:

Capacity:

Increment:

Frequency:

Next Due:

As Found

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23460	23460	0
33460	33460	0
44300	44300	0

Repeatability

44300	44300	0
-------	-------	---

Off-centre Loading

6820	6820	0
------	------	---

Engineer Name:

Definitive

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23460	23460	0
33460	33460	0
44300	44300	0

Repeatability

44300	44300	0
-------	-------	---

Off-centre Loading

6820	6820	0
------	------	---

Trading Standards Maximum Tolerance

+/-

Traceable Weights Used:



West End Grange
 Epworth Water Gardens
 West End Road
 Epworth
 Doncaster
 DN9 1LA



Certificate No:

50019

Tel: 01427 875 867
 Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name:

Address:

Postcode:

Contact Name:

Test Date:

Location:

Manufacturer:

Serial/Plant No:

Capacity:

Increment:

Frequency:

Next Due:

As Found

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23460	23460	0
33460	33460	0
44300	44300	0

Repeatability

44300	44300	0
-------	-------	---

Off-centre Loading

6820	6820	0
------	------	---

Engineer Name:

Definitive

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23460	23460	0
33460	33460	0
44300	44300	0

Repeatability

44300	44300	0
-------	-------	---

Off-centre Loading

6820	6820	0
------	------	---

Trading Standards Maximum Tolerance

+/-

Traceable Weights Used: