



West End Grange
 Epworth Water Gardens
 West End Road
 Epworth
 Doncaster
 DN9 1LA



Certificate No:

52419

Tel: 01427 875 867
 Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name:

Address:

Postcode:

Contact Name:

Test Date:

Location:

Manufacturer:

Serial/Plant No:

Capacity:

Increment:

Frequency:

Next Due:

As Found

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	15980	20
23520	23500	20
33520	33480	40
44360	44300	60

Repeatability

44360	44300	60
-------	-------	----

Off-centre Loading

6820	6820	0
------	------	---

Engineer Name:

Definitive

Measurement Points	Reading	Deviation
0	0	0
8000	8000	0
16000	16000	0
23520	23520	0
33520	33520	0
44360	44360	0

Repeatability

44360	44360	0
-------	-------	---

Off-centre Loading

6820	6820	0
------	------	---

Trading Standards Maximum Tolerance

+/-

Traceable Weights Used: